

User Manual

Version 2.0

Product name: 2X HoTaq Real-time PCR Kit (Regular ROX)

Cat #: HTP400

Description:

Contains regular ROX. For ABI 7000, 7300, 7700, 7900 machines.

This product enables sensitive detection of DNA and fast thermocycling, using proprietary hot-start PCR technology developed at MCLAB (patent pending).

Storage:

Store at -20°C . To avoid repeated freeze-thaw, opened vials should be kept at 4°C .

Primer and probe design:

To achieve the best performance, appropriate software, such as ABI's Primer ExpressTM, should be used.

1. T_m: 60°C for primers and $68\sim 70^{\circ}\text{C}$ for probes
2. Amplicon size should be small, $<150\text{bp}$
3. To avoid secondary structures and avoid more than 3 consecutive Gs in primers and probes
4. Primers should be 17 ~ 30 nucleotides in length and should not have complementary 3' ends

Reaction conditions:

95°C , 10 min. => (95°C , 5 sec. => 60°C , 30 sec.) for 50 cycles.

Tips for good performance

To achieve accurate quantification, it is highly recommended to do replicates. Three is the minimal number of replicates to obtain a standard deviation. It is important to reduce pipetting errors. Three ways to minimize pipetting errors are:

1. To prepare an amplicon specific master mix that includes PCR reaction mix, primers, and probes.
2. To use a repeat pipet.
3. To pipet volumes within the manufacture's suggested range.